



Product Service

EC Certificate

Full Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)
(Devices in Class IIa, IIb or III)

No. G1 12 04 15198 023

Manufacturer: Bionic Medizintechnik GmbH

Max-Planck-Straße 21
61381 Friedrichsdorf
GERMANY

Product Category(ies): **Fistula Needles**
Dialysis Catheters
Connection and Disconnection Sets
(Products see attachment)

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for design, manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex II. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class III devices an additional Annex II (4) certificate is mandatory. See also notes overleaf.

Report No.: 713000060

Valid from: 2012-05-29

Valid until: 2017-05-28

Date, 2012-04-18

Hans-Heiner Junker



TÜV SÜD Product Service GmbH is Notified Body with identification no. 0123

Page 1 of 3



Product Service

EC Certificate**Full Quality Assurance System**Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)
(Devices in Class IIa, IIb or III)

No. G1 12 04 15198 023

Facility(ies):Bionic Medizintechnik GmbH
Max-Planck-Straße 21, 61381 Friedrichsdorf, GERMANY

Attachment to Certificate no G1 12 04 15198 023
dated 2012-04-18



Product Service

Products:

Fistula needles, fixed wing
Fistula needles, rotating hub
Fistula needles, „Single Needle“
Fistula needles, „Button Hole“
Safety fistula needles
Dialysis Connection and Disconnection Sets
Dialysis catheters, single and multi lumen
Dialysis catheter sets
Dialysis catheter, accessories
DEMERS catheter, implantable

Munich, CRT2, 2012-04-18

A handwritten signature in black ink, appearing to read 'H.-H. Junker'.

Hans-Heiner Junker

